



ANALYSIS REPORT

Page 1 of 2

Client:	Hawkes Bay Regional Council	Lab No:	1446345	SPV1
Contact:	V Lyon	Date Registered:	03-Jul-2015	
	C/- Hawkes Bay Regional Council	Date Reported:	09-Jul-2015	
	Private Bag 6006	Quote No:	69783	
	NAPIER 4142	Order No:	N37013	
		Client Reference:	Whangawehi	
		Add. Client Ref:	312-302	
		Submitted By:	V Lyon	

Sample Type: Aqueous

Sample Name:	56373 - Coops Crk US Whangawehi - 3306	56374 - Reserve Stream - Trib of Whangawehi - 3307	56375 - Whangawehi US Reserve Confl - 3308		
Lab Number:	1446345.1	1446345.2	1446345.3		
Individual Tests					
Escherichia coli	cfu / 100mL	190 #1	43	200	-
HBRC Standard River					
Total Suspended Solids	g/m ³	3.0	1.7	1.6	-
Total Nitrogen	g/m ³	0.25	0.23	0.42	-
Total Ammoniacal-N	g/m ³	< 0.005	< 0.005	0.006	-
Nitrite-N	g/m ³	0.0020	< 0.0010	< 0.0010	-
Nitrate-N	g/m ³	0.055	0.114	0.114	-
Nitrate-N + Nitrite-N	g/m ³	0.057	0.115	0.115	-
Total Kjeldahl Nitrogen (TKN)	g/m ³	0.19	0.12	0.30	-
Dissolved Reactive Phosphorus	g/m ³	0.0174	0.084	0.115	-
Total Phosphorus	g/m ³	0.024	0.111	0.143	-

Analyst's Comments

Please interpret these microbiological results with caution as the age of the sample is unknown. Samples must be less than 24 hours old at the time of testing. Please record both the time and date of sampling on your submission form and the sample bottles.

#1 Statistically estimated count based on the theoretical countable range for the stated method.

SUMMARY OF METHODS

The following table(s) gives a brief description of the methods used to conduct the analyses for this job. The detection limits given below are those attainable in a relatively clean matrix. Detection limits may be higher for individual samples should insufficient sample be available, or if the matrix requires that dilutions be performed during analysis.

Sample Type: Aqueous

Test	Method Description	Default Detection Limit	Sample No
HBRC Standard River		0.0010 - 0.5 g/m ³	1-3
Filtration, Unpreserved	Sample filtration through 0.45µm membrane filter.	-	1-3
Total Kjeldahl Digestion	Sulphuric acid digestion with copper sulphate catalyst.	-	1-3
Total Phosphorus Digestion	Acid persulphate digestion.	-	1-3
Total Suspended Solids	Filtration of a 2L sample using Whatman 934 AH, Advantec GC-50 or equivalent filters (nominal pore size 1.2 - 1.5µm), gravimetric determination. APHA 2540 D 22 nd ed. 2012.	0.5 g/m ³	1-3
Total Nitrogen	Calculation: TKN + Nitrate-N + Nitrite-N. Please note: The Default Detection Limit of 0.05 g/m ³ is only attainable when the TKN has been determined using a trace method utilising duplicate analyses. In cases where the Detection Limit for TKN is 0.10 g/m ³ , the Default Detection Limit for Total Nitrogen will be 0.11 g/m ³ .	0.05 g/m ³	1-3



This Laboratory is accredited by International Accreditation New Zealand (IANZ), which represents New Zealand in the International Laboratory Accreditation Cooperation (ILAC). Through the ILAC Mutual Recognition Arrangement (ILAC-MRA) this accreditation is internationally recognised. The tests reported herein have been performed in accordance with the terms of accreditation, with the exception of tests marked *, which are not accredited.

Sample Type: Aqueous			
Test	Method Description	Default Detection Limit	Sample No
Total Ammoniacal-N Trace	Phenol/hypochlorite colorimetry. Flow injection analyser. (NH ₄ -N = NH ₄ ⁺ -N + NH ₃ -N). APHA 4500-NH ₃ H 22 nd ed. 2012.	0.005 g/m ³	1-3
Nitrite-N Trace	Automated Azo dye colorimetry, Flow injection analyser. APHA 4500-NO ₃ ⁻ I 22 nd ed. 2012 (modified).	0.0010 g/m ³	1-3
Nitrate-N	Calculation: (Nitrate-N + Nitrite-N) - NO ₂ N. In-House.	0.0010 g/m ³	1-3
Nitrate-N + Nitrite-N Trace	Total oxidised nitrogen. Automated cadmium reduction, flow injection analyser. APHA 4500-NO ₃ ⁻ I 22 nd ed. 2012 (modified).	0.0010 g/m ³	1-3
Total Kjeldahl Nitrogen (TKN)	Total Kjeldahl digestion, phenol/hypochlorite colorimetry. Discrete Analyser. APHA 4500-N _{org} D. (modified) 4500 NH ₃ F (modified) 22 nd ed. 2012.	0.10 g/m ³	1-3
Dissolved Reactive Phosphorus (trace)	Filtered sample. Molybdenum blue colorimetry. Flow injection analyser. APHA 4500-P G 22 nd ed. 2012.	0.0010 g/m ³	1-3
Total Phosphorus	Total phosphorus digestion, ascorbic acid colorimetry. Discrete Analyser. APHA 4500-P B & E (modified from manual analysis) 22 nd ed. 2012. Also modified to include the use of a reductant to eliminate interference from arsenic present in the sample. NWASCA, Water & soil Miscellaneous Publication No. 38, 1982.	0.004 g/m ³	1-3
Escherichia coli	Membrane filtration, Count on mFC agar, Incubated at 44.5°C for 22 hours, MUG Confirmation. Analysed at Hill Laboratories - Microbiology; 1 Clow Place, Hamilton. APHA 9222 G, 22 nd ed. 2012.	1 cfu / 100mL	1-3

These samples were collected by yourselves (or your agent) and analysed as received at the laboratory.

Samples are held at the laboratory after reporting for a length of time depending on the preservation used and the stability of the analytes being tested. Once the storage period is completed the samples are discarded unless otherwise advised by the client.

This report must not be reproduced, except in full, without the written consent of the signatory.

Ara Heron BSc (Tech)
Client Services Manager - Environmental Division